

BIOMEDICAL RESEARCH SERVICE CENTER

UNIVERSITY at BUFFALO, STATE UNIVERSITY of NEW YORK

Department of Biochemistry, Attn: Dr. Lee, University at Buffalo, 3435 Main Street, Buffalo, NY 14214, USA
Tel/Fax: (716) 829-3106 Email: chunglee@buffalo.edu Web: www.bmrservice.com

Glucose-6-phosphate Dehydrogenase (G6PD) Assay Kit (Cat #: E-110)

COMPONENTS: G6PD Assay Solution- 10 ml (for 200 assays); **store in aliquots at -80°C after first thawing**
10x Cell Lysis Solution- 25 ml, store at 4°C

PRODUCT DESCRIPTION: The G6PD assay is based on the reduction of the tetrazolium salt INT in a NADH-coupled reaction to INT-formazan, which exhibits an absorption maximum at 492 nm (molar extinction coefficient = $18 \text{ mM}^{-1} \text{ cm}^{-1}$) and allows for sensitive detection of G6PD activity in serum and tissue samples. Assay solution is stable for several years if stored and handled properly.

Plasma: Plasma samples may need to be diluted with ice-cold 1x Cell Lysis Solution (from 10-fold dilution of 10x Cell Lysis Solution with dH_2O) to obtain assay linearity. Samples should be store at -80°C.

Preparation of cell/tissue extracts:

1. Wash $10^5 - 10^6$ cells with ice-cold phosphate-buffered saline (PBS). Animal tissue should be washed with PBS thoroughly to remove blood cells. Freeze cell pellet and tissue at -80°C until use.
2. Add 50 – 100 μl ice-cold 1x Cell Lysis Solution to cell pellet. Extract cells by pipetting up and down (gently but thoroughly). Leave lysate on ice for 5 min with intermittent gentle agitation. Centrifuge lysate in a refrigerated microfuge for 3 min at maximum speed ($\sim 13,000$ rpm) and harvest supernatant for G6PD assay. For tissue extraction, weigh ~ 50 mg tissue and homogenize in 1 ml ice-cold 1x Cell Lysis Solution. Centrifuge homogenate at 4°C for 3 min at maximum speed and harvest supernatant. Store cell lysate and tissue homogenate at -80°C.
3. Perform protein assay to determine sample protein concentration. Equalize sample protein concentration by diluting with ice-cold 1x Cell Lysis Solution. A suggested protein concentration range is 0.2 – 2 mg/ml.

* Single time point assay:

1. Thaw G6PD Assay Solution and keep on ice. Gently agitate solution prior to pipetting. Add 10 μl sample to a 96-well plate. Use 10 μl 1x Cell Lysis Solution as blank. Reaction is initiated by adding 50 μl G6PD Assay Solution to each well. Mix contents by brief gentle agitation. Cover plate and incubate at 37°C for 30 min or 60 min (for more intense reaction).
2. Stop reaction by adding 50 μl 3% Acetic acid (not included in the kit) to each well followed by brief gentle agitation. Measure $\text{O.D.}_{492 \text{ nm}}$ using a plate reader. Subtract blank reading from sample reading. G6PD activity in IU/L unit = $\mu\text{mol}/(\text{L}\cdot\text{min}) = \text{O.D.} \times 1000 \times 110 \mu\text{l} / (30 \text{ min} \times 0.6 \text{ cm} \times 18 \times 10 \mu\text{l}) = \text{O.D.} \times 33.95$. If incubation for 60 min, G6PD activity in IU/L unit = **O.D. x 16.98**

**** Dual time point assay:** Use this format to eliminate interference that may be caused by sample hemolysis.

1. Set up assay in duplicate 96-well plates for two selected time points. We suggest using 20 and 40 min initially. Longer assay time can be adopted if weak G6PD activity is observed (e.g. 30 & 50 min or 40 & 60 min). For samples exhibiting high G6PD activity, select two time points with a 10-min interval (e.g. 10 & 20 min or 20 & 30 min). Optimal assay time may be determined empirically.
2. For the 20/40 min format, add 10 μl sample to each of the two 96-well plates. Include a 10 μl blank on each plate. Add 50 μl G6PD Assay Solution to each well and mix by brief gentle agitation. Cover plates and incubate at 37°C for 20 min. Remove one plate and stop reaction by adding 50 μl 3% Acetic acid. Remove the second plate at 40 min and stop reaction by 50 μl Acetic acid. Measure $\text{O.D.}_{492 \text{ nm}}$. Subtract blank reading from sample reading. G6PD activity in IU/L unit = $(\text{O.D.}_{40 \text{ min}} - \text{O.D.}_{20 \text{ min}}) \times 1000 \times 110 / (20 \times 0.6 \times 18 \times 10) = \Delta\text{O.D.} \times 50.93$. If using a 10-min interval, G6PD activity in IU/L unit = **$\Delta\text{O.D.} \times 101.85$**

Additional information:

- Interference caused by sample hemolysis can also be eliminated by including a sample control well containing 10 μl sample, 50 μl dH_2O and 50 μl 3% Acetic acid. The control well reading is then subtracted from sample reaction reading.
- Note that “0.6 cm” in the equation is the typical light path in a 96-well plate and may be custom adjusted as needed.
- A 3% Acetic acid solution needs to be prepared for reaction termination. The assay solution contains DMSO and iodinitrotetrazolium violet. Please refer to the product page of our website or contact us for MSDS information.